



LAB N° 0032 L

LARISA FACE COVER S.A.  
6th KLM NATIONAL ROAD LARISSA  
ATHENS, 41110  
GR

Client Account Number: A00895229RF5  
Eurofins Quote Number: PO9YPH20081401

**Eurofins Sample Number LV20AB9066-1**

<b>Original Received Date:</b>	24-Nov-2020
<b>Description:</b>	§ 3ply non woven disposable surgical masks
<b>Lot Number:</b>	§ L202891

Analysis	Result	Unit
<b># In vitro cytotoxicity</b>		
RESULT:	Not cytotoxic	----
Test item:	30mm <sup>2</sup>	----
Vehicle:	Routine medium	----
Positive control:	30mm <sup>2</sup> of Latex	----
Negative control:	30mm <sup>2</sup> of HDPE	----
Notes:	N/A	----

**Addendum #1: Qualitative and quantitative evaluation**

Method: ISO 10993-5:2009

Analysis Date: 02-Dec-2020 to 04-Dec-2020

**Supplemental Information**

For LV20AB9066-1:

Sample description: green surgical mask

Storage: room temperature


# # The Eurofins Testing Laboratory is accredited by ACCREDIA only for tests marked with #. This test report may not be reproduced in part unless expressly approved in writing by Eurofins Biolab S.r.l. The test results relate only to the tested items. Sampling, except specific indication on test report, is always intended to be made by the Sponsor. Information provided by the Sponsor are under Sponsor responsibility, and are marked by the symbol "§".



LAB N° 0032 L

**Contracted Company: Eurofins Biolab Srl (Vimodrone)**Via Bruno Buozzi, 2, Vimodrone, MI 20055 Italy  
InfoFarma@eurofins.com

*This Certificate of Analysis cannot be modified and/or partially reproduced without the written authorization of the laboratory.  
Questions about this report should be directed to your project manager or the general email listed above.*

 <b>eurofins</b>	In vitro cytotoxicity ISO10993-5:2009	1-P-QM-TEM-9070229
	Test by Direct Contact - NRU	Addendum N. 1

Study ID: //

Sample ID: LV20AB9066-1

Test start: 02/12/2020

Test end: 04/12/2020

Cell line	Manufacturer	
Mammal fibroblasts BALB/3T3 clone A31 (ATCC® CCL163™)	American Type Culture Collection (ATCC)	
Reagent	Manufacturer	Batch
Dulbecco's Modified Eagle Medium (DMEM)	Sigma-Aldrich	RNBj2369
Fetal Bovine Serum (FBS)	Sigma-Aldrich	17H114
Penicillin/Streptomycin solution	Sigma-Aldrich	0000088375

### QUALITATIVE EVALUATION

Contact time: 24 h					Contact time: 24 h				
PLATE 1					PLATE 2				
	Blank	Vehicle control	Negative control	Test Sample		Blank	Vehicle control	Positive control	Test Sample
Replicate 1		0	0	0	Replicate 1		0	4	0
Replicate 2		0	0	0	Replicate 2		0	4	0
Replicate 3		0	0	0	Replicate 3		0	4	0

### INTERPRETATION OF RESULT

Grade	Reactivity	Conditions of all Cultures
0	None	No detectable zone around or under specimen
1	Slight	Some malformed or degenerated cells under specimen
2	Mild	Zone limited to area under specimen
3	Moderate	Zone extending specimen size up to 1.0 cm
4	Severe	Zone extending farther than 1.0 cm beyond specimen


### ACCEPTANCE CRITERIA

Negative control	
Grade ≤1	VALID

Positive control	
Grade ≥3	VALID

### RESULTS

	Reactivity grade
Test Sample	0

 <b>eurofins</b>	In vitro cytotoxicity ISO10993-5:2009	1-P-QM-TEM-9070229
	Test by Direct Contact - NRU	Addendum N. 1

Study ID: //

Sample ID: LV20AB9066-1

Test start: 02/12/2020

Test end: 04/12/2020

**QUANTITATIVE EVALUATION - 540 nm**

Contact time: 24 h					Contact time: 24 h				
PLATE 1					PLATE 2				
	Blank	Vehicle control	Negative control	Test Sample		Blank	Vehicle control	Positive control	Test Sample
Replicate 1	0,048	1,496	1,463	1,486	Replicate 1	0,047	1,464	0,046	1,424
Replicate 2	0,051	1,442	1,217	1,454	Replicate 2	0,048	1,434	0,040	1,442
Replicate 3	0,046	1,501	1,491	1,479	Replicate 3	0,045	1,458	0,042	1,433

**MEAN, STANDARD DEVIATION, CV% AND VIABILITY**

	Mean OD	Standard Deviation	CV %	Mean OD - Mean OD Blanks	Viability %
Blanks	0,048				
Vehicle control	1,466	0,028	1,879	1,418	100
Negative control	1,390	0,151	10,844	1,343	95
Positive control	0,043	0,003	7,160	-0,005	0
Test Sample	1,453	0,025	1,721	1,406	99

**ACCEPTANCE CRITERIA**

	Vehicle $\geq 0,3$			
Mean OD	VALID			
	Negative control $\geq 70\%$	Positive Control $< 70\%$		
Quantitative Evaluation	VALID	VALID		
	Vehicle control	Negative control	Positive control	Test Sample
CV between replicates $\leq 15\%$	VALID	VALID	VALID	VALID

$$\% \text{ viability} = \frac{\text{OD}_{\text{mean sample}} - \text{OD}_{\text{mean blank}}}{\text{OD}_{\text{mean vehicle}} - \text{OD}_{\text{mean blank}}} \times 100$$

**INTERPRETATION OF RESULTS**

Reduction of Viability	Result
$\leq 30\%$	Not Cytotoxic
$> 30\%$	Cytotoxic

**RESULTS**

	Reduction Viability %	
Test Sample	1	NOT CYTOTOXIC

\*0 (% viability test item  $\geq$  vehicle); 100 (% viability test item  $\leq$  Blank)


Finished on: 04/12/20

Technician signature: JN

Date: 04/12/20

Approved by: EP

Rev.8	Local Reference: Mod. PS/TOX/020.D	Page 2 of 2
©This document is copyright of Eurofins Scientific Group	Documento completo approvato in ETQ	

	<b>Medical Device Testing</b>	<b>Test Facility</b> Eurofins Biolab S.r.l.	<b>Sample ID:</b> Page: 1 of 1
---	-------------------------------	--	-----------------------------------

**ADDENDUM N.2: EXPERIMENTAL REPORT**

REFERENCE/GUIDELINE:	- ISO 10993-5:2009 - In vitro cytotoxicity					
CELL LINE	Mammal fibroblasts BALB/3T3 clone A31 (ATCC®; CCL163™) Source: ATCC.					
MATERIALS	Foetal Bovine Serum (FBS), Trypan Blue, Neutral Red dye, SDS, Penicillin/Streptomycin solution, Dulbecco's Modification of Eagle's Medium (DMEM), Dulbecco's Phosphate buffer solution (DPBS)				Sigma-Aldrich	
	Trypsin-EDTA				Lonza	
	Acetic Acid, Ethanol				VWR	
	Water for Injection				Eurospital	
	High density polyethylene (HDPE, USP Reference Standard negative control)				Nova Chimica (KOM357)	
	Latex from laboratory gloves				Artsana (1811191C170)	
EQUIPMENT	Laminar flow hood, CO <sub>2</sub> incubator, Microplate reader Mod EL800, Chronometer, Common laboratory equipment, Water, Inverted Microscope Diavert,Orbital shaker, Refrigerator					
EXPERIMENTAL DESIGN						
The experimental design included two 12-well plate containing a subconfluent cell monolayer subdivided in the following groups:						
GROUPS	REPLICATES PLATE N.1			REPLICATES PLATE N.2		
	Blank	Blank	Blank	Blank	Blank	Blank
	Vehicle	Vehicle	Vehicle	Vehicle	Vehicle	Vehicle
	Negative control	Negative control	Negative control	Positive control	Positive control	Positive control
	Test sample	Test sample	Test sample	Test sample	Test sample	Test sample
BLANK		Supplemented culture medium alone (without cells).				
VEHICLE		Supplemented culture medium (without test sample).				
TEST SAMPLE		30 mm <sup>2</sup> of the test sample were placed in the middle of each well.				
NEGATIVE CONTROL		The negative control was represented by 30 mm <sup>2</sup> of HDPE placed in the middle of each well.				
POSITIVE CONTROL		The positive control was represented 30 mm <sup>2</sup> of latex placed in the middle of each well.				
TREATMENT: Verified that a subconfluent monolayer was present, supplemented culture medium was replaced and the test sample was added. The plates were incubated in a thermostat at 37°C ±1°C in a 5% CO <sub>2</sub> atmosphere for 24 hours. This procedure was repeated for positive and negative controls.						
QUALITATIVE EVALUATION (GRADE OF CYTOTOXICITY): After 24 hours the plates were observed under an inverted microscope and biological reactions were evaluated following a 0 to 4 scale according to ISO10993-5:2009.						
QUANTITATIVE EVALUATION (OPTICAL DENSITY): After microscopic observation, cells were treated with Neutral Red Medium for 3 hours at 37°C ±1°C in 5% CO <sub>2</sub> atmosphere. Subsequently, the Neutral Red medium was removed and each well was rinsed with DPBS. The plates were totally made dry reversing the plates, then Desorb Solution was added and the plates were incubated for 10 minutes at room temperature with gentle agitation to form a homogeneous solution. Optical density was measured at 540nm by Gen5 software (Biotek) using microtiter plate reader.						
$\% \text{ of cell viability} = \frac{\text{OD test sample} - \text{OD blank}}{\text{OD vehicle} - \text{ODblank}} \cdot 100$						
ACCEPTABILITY CRITERIA	QUALITATIVE EVALUATION			Negative control ≤ 1; Positive control ≥ 3		
	QUANTITATIVE EVALUATION			The OD mean of the vehicle must be ≥ 0,3. The positive control % cellular viability must be < 70%. The negative control % cellular viability must be ≥ 70%. Coefficient of variation of each group must be ≤15%.		
INTERPRETATION OF RESULTS	The achievement of a numerical grade greater than 2 is considered a cytotoxic effect. A cellular viability reduction more than 30% is considered a cytotoxic effect.					
QUALITY CRITERIA	Satisfied					
QUANTITATIVE EVALUATION	Cells treated with test sample have shown a cell viability reduction of 1±0,76%.					

This test report may not be reproduced in part unless expressly approved in writing by Eurofins Biolab S.r.l. The test results relate only to the tested items. Sampling, except specific indication on test report, is always intended to be made by the Sponsor. Information provided by the Sponsor are under Sponsor responsibility.

**Eurofins Biolab S.r.l.**

Società con Socio unico sottoposta  
a direzione e coordinamento della società  
Eurofins Pharma Services Italia Holding  
parte di Eurofins Scientific Group  
[InfoFarma@eurofins.com](mailto:InfoFarma@eurofins.com)  
[www.eurofins.it](http://www.eurofins.it) [www.biolab.it](http://www.biolab.it)

**Via Bruno Buozzi, 2**  
**20090 Vimodrone (MI)**  
**Tel. + 39-022507151**  
**Fax + 39-0225071599**

[certificata@pec.biolab.it](mailto:certificata@pec.biolab.it)

C.SOC. € 100.000 i.v.  
P. IVA 00762140960  
C.F. 03765750157  
REA MI 966696  
D-U-N-S 429117112  
CIT005